

## **Nordic Immunohistochemical Quality Control**

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

## Recommended protocol for TdT

## Obtained in run 52

08 Dec 2017

**Immunostainer** 

Type: Intellipath Biocare medical

**Primary antibody** 

Clone: SEN28

Producer: Leica/novocastra

Product no. / lot no.: NCL-L-TdT-339 / 6039147

Diluent: Davinci green

Dilution factor: 1:50

Incubation time / temperature: 30 min. / 20°C

Epitope retrieval, HIER

Device: Pressure Cooker
Buffer: Biocare medical Borg

Heating time at max. temp.: 15 min. Maximum heating temp.: 110°C

Visualization system

Producer: Biocare medicare

Product / no: Mach4 universal HRP polymer / BRI 4012L

Linker: None Incubation time polymer: 10 min. Incubation temperature: 20°C

Chromogen

Producer: Biocare medical

Product / no: Mach4 universal DAB / BRI 4012L

Incubation time / temperature: 5 min. / 20°C

Enhancement: None

## Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.