

Nordic Immunohistochemical Quality Control

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

Recommended protocol for IgK

Obtained in run

21 Aug 2016

Immunostainer

Type: TechMate, Dako

Primary antibody

Clone: Polyclonal

Producer: Dako

Product no. / lot no.: A0191 / NS

Diluent: Antibody Diluent

Dilution factor: 1:8000

Incubation time / temperature: 25 min. / 20°C

Epitope retrieval, HIER

Device: Microwave oven
Buffer: Dako TRS Low pH

Heating time at max. temp.: 20 min. Maximum heating temp.: 100°C

Visualization system

Producer: Dako

Product / no: REAL EnVision Rabbit/Mouse / K5007 / K4065

Incubation time polymer: 25 min. Incubation temperature: 20°C

Chromogen

Producer: Dako

Product / no: DAB+ / K5007 / K4065

Incubation time / temperature: 15 min. / 20°C

Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.