

Nordic Immunohistochemical Quality Control

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

Recommended protocol for CD23

Obtained in run 50

03 Mar 2017

Immunostainer

Type: Dako Omnis

Primary antibody

Clone: SP23

Producer: Spring Bioscience
Product no. / lot no.: M3234 / 150916LVG
Diluent: Antibody Diluent

Dilution factor: 1:50

Incubation time / temperature: 20 min. / 32°C

Epitope retrieval, HIER

Device: On Board / On Machine

Buffer: Dako Omnis Target Retrieval Solution, Low pH

Heating time at max. temp.: 30 min. Maximum heating temp.: 97°C

Visualization system

Producer: Dako Omnis

Product / no: EnVision Flex / GV800/GV823

Linker: None Incubation time polymer: 20 min. Incubation temperature: 32°C

Chromogen

Producer: Dako Omnis

Product / no: DAB+ Substrate Chromogen System / GV825

Incubation time / temperature: 5 min. / 32°C

Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.