


### Purpose

The primary focus of this assessment is evaluation of the technical performance of HER2 Brightfield in-situ hybridization (BRISH) tests performed by the NordiQC participants for demonstration and establishment of the HER2 gene amplification level in breast carcinomas. In addition, the participants are asked to interpret and score the amplification status in the breast carcinomas and submit these to NordiQC in order to evaluate the inter-observer variability. The evaluation of inter-observer concordance is applicable for participants using either BRISH based tests or Fluorescent in-situ hybridization (FISH) based tests. The obtained assessment marks in NordiQC is indicative of the performance of the tests but due to the limited number and composition of samples, internal validation and extended quality control, e.g. regularly measuring the HER2 results, is necessary.

### Material

Table 1. Content of the multi-block used for the NordiQC HER2 ISH assessment, run H27

	HER2 IHC*	Dual - BRISH**	Dual - BRISH**	FISH***	FISH***
	IHC score	HER2/chr17 ratio $\square$	HER2 copies	HER2/chr17 ratio $\square$	HER2 copies
1. Breast carcinoma	2+	3	6.9	2.44	6.1
2. Breast carcinoma	2+	1.66	3.65	1.19	2.15
3. Breast carcinoma	1+	1.55	2.95	1.33	2.65
4. Breast carcinoma	3+	7.96	9.95	6.93	11.45
5. Breast carcinoma	0	1.03	1.95	0.93	1.9

\* PATHWAY<sup>®</sup> (Ventana), data from two reference labs.

\*\* Ventana HER2 Dual ISH DNA Probe Cocktail, data from one reference lab.

\*\*\* HER2 FISH (Zytovision), data from one reference lab.

$\square$  HER2/chr17: HER2 gene/chromosome 17 ratios.

All tissues were fixed for 24-48 hours in 10% neutral buffered formalin according to the ASCO/CAP 2023 guidelines for tissue preparation of breast tissue for HER2 ISH analysis.

### HER2 BRISH, Technical assessment

The NordiQC assessors evaluate the technical quality of the BRISH tests and at this point do not conduct a precise estimation of the HER2 amplification status. The main criteria for the technical evaluation are as listed below.

Staining was assessed as **optimal**, if the HER2/chr17 ratios could be evaluated in all five tissues and no technical artefacts compromising the interpretation being observed. Small blank spots <25% of the core was accepted.

Staining was assessed as **good**, if the HER2/chr17 ratios could be evaluated in all five tissues, but the interpretation was slightly compromised e.g. due to excessive retrieval, weak or excessive counterstaining or large negative areas with no signals (>25% of the core)

Staining was assessed as **borderline** if one of the tissues could not be evaluated properly e.g. due to weak or missing signals, a low signal-to-noise ratio, excessive background staining or impaired morphology.

Staining was assessed as **poor** if two or more of the tissue cores could not be evaluated properly e.g. due to weak or missing signals, a low signal-to-noise ratio, excessive background staining or impaired morphology

Note that the assessment criteria were modified in run H24 compared to previous assessments. Large negative areas of >25% of the individual tissue cores were accepted providing the HER2 gene amplification level still reliably could be evaluated. However, a slide with large negative areas was not compatible with an optimal assessment mark and was downgraded to good providing an otherwise optimal result being observed.

## HER2 BRISH and FISH interpretation

For both BRISH and FISH, participating laboratories were asked to submit a scoring sheet with their interpretation of the HER2/chr17 gene status. Results were compared to NordiQC FISH and BRISH data from reference laboratories to analyze scoring consensus.

Consensus scores from the NordiQC BRISH/FISH reference laboratories

- Breast carcinoma, no. 2, 3 and 5: non-amplified
- Breast carcinoma, no. 1 and 4: amplified

*The ASCO/CAP 2023 guidelines were applied for the interpretation of the HER2 status:*

**Amplified:** HER2/chr17 ratio  $\geq 2.0$  using a dual probe assay with an average  $\geq 4$  HER2 copies per cell/nucleus. Using a single probe assay an average of  $\geq 6$  HER2 copies per cell/nucleus. (Group 1)

**Equivocal** (Additional work-up required):

HER2/chr17 ratio of  $\geq 2.0$  using a dual probe assay with an average of  $< 4$  HER2 gene copies per cell/nucleus (Group 2)

HER2/chr17 ratio of  $< 2.0$  using a dual probe assay with an average of  $\geq 6$  HER2 gene copies per cell/nucleus (Group 3)

HER2/chr17 ratio of  $< 2.0$  using a dual probe assay with an average of  $\geq 4$  and  $< 6$  HER2 gene copies per cell/nucleus (both dual and single probe assay) (Group 4)

**Unamplified:** HER2/chr17 ratio  $< 2.0$  using a dual probe assay with an average  $< 4$  HER2 gene copies per cell/nucleus (both dual and single probe assay) (Group 5)

## Participation


Number of laboratories registered for HER2 BRISH	185
Number of laboratories returning slides	174 (94%)
Number of laboratories returning scoring sheet	147 (85%)
Number of laboratories registered for HER2 FISH	76
Number of laboratories returning scoring sheet	70 (92%)

At the date of the technical assessment, 94% of the participants had returned the circulated NordiQC slides. All slides returned after the assessment were assessed and laboratories received advice if the result was insufficient, but the data were not included in this report.

## Performance history

In this assessment run H27 the overall pass rate of 85% was obtained and improved compared to the level gained in the previous assessment run H26 (78% pass rate), and further improved compared to the levels obtained in the previous assessment runs as illustrated in Graph 1. The improvement seen in the latest runs is mainly caused by new modified assessment criteria applied in run H24 allowing large negative areas of  $>25\%$  in one or more of the tissue cores providing an evaluation of the HER2/chr17 ratio still adequately could be obtained.

Graph 1. **Proportion of sufficient results for HER2 BRISH in NordiQC assessments, 2019 – 2025**



## Results BRISH, technical assessment

In total, 174 laboratories participated in this assessment. 148 laboratories (85%) achieved a sufficient mark (optimal or good). Results are summarized in Table 2.

Table 2. HER2 BRISH systems and assessment marks for BRISH HER2 run H27.

Two colour HER2 systems	n	Vendor	Optimal	Good	Borderline	Poor	Suff. <sup>1</sup>	OR <sup>2</sup>
INFORM™ HER2 Dual ISH 780-4422 / 800-4422	1	Ventana/Roche	0	0	0	1	-	-
VENTANA HER2 Dual ISH 800-6043	152	Ventana/Roche	71	62	14	5	88%	47%
VENTANA HER2 Dual ISH + IHC 800-6043 + HER2 IHC (GPA*)	16	Ventana/Roche	6	5	4	1	69%	38%
ZytoDot® 2C C-3022 / C-3032	4	ZytoVision	2	2	0	0	-	-
One colour HER2 systems								
ZytoDot® C-3003	1	ZytoVision	0	0	1	0		
Total	174		79	69	19	7		
Proportion			45%	40%	11%	4%	85%	

1) Proportion of Sufficient Results (≥5 assessed protocols).

2) Proportion of Optimal Results (≥5 assessed protocols).

\* GPA; Gene Protein Assay (HER2 BRISH + PATHWAY HER2 IHC).

## Comments

In this run and in concordance with the latest assessments, the vast majority of participants (97%) used BRISH HER2 systems from Ventana/Roche. 97% (168 of 174 participants) used the VENTANA HER2 Dual ISH DNA Probe Cocktail (800-6043) and 0,5% (1 of 174) the INFORM™ HER2 Dual ISH assay (780-4422/800-4422). 2,5% (5 of 174) used the HER2 BRISH systems ZytoDot®.

9% of participants (16 of 174) used the VENTANA HER2 Dual ISH DNA Probe Cocktail (800-6043) in combination with HER2 IHC providing a Gene Protein Assay (GPA). In the evaluation of the technical assessment, only the HER2 BRISH results were addressed.

As shown in Table 2, a technically optimal performance for the demonstration of HER2/Chr17 signals permitting an adequate evaluation of the HER2 gene amplification status in all the five breast carcinomas included in the multi-tissue block was obtained by both Ventana/Roche dual-colour BRISH systems and the ZytoVision ZytoDot® 2C system.

The insufficient results were most frequently characterized by large negative areas in one or more of the breast carcinoma samples compromising the evaluation of HER2/chr 17 ratio, but also caused by impaired morphology, generally weak or missing signals for either HER2 and/or chr17.

In this assessment 77% (20 of 26) of the insufficient results were characterized by impaired morphology, or negative areas (>25%) in one or more tissue cores either as single feature or combined with other artefacts as false negative signals for HER2/chr 17 and/or weak counterstaining. In the remaining 23% of the insufficient results these were mainly caused by, weak staining and focal negative areas.

As described in the assessment report for run H23 (2023) and illustrated in Graph 1, no significant improvement in pass rates had been obtained for HER2 BRISH in the period from 2019-2023 and a cumulated average level of 65% was obtained in the NordiQC assessment runs H15-H23. In all these runs, the ISH rejection criteria defined in the 2013/2018 ASCO/CAP HER2 guidelines were applied. In brief, repeated test must be performed if more than 25% of the signals/cells cannot be interpreted in the sample evaluated. However, by internal discussions within the NordiQC assessor panel and from correspondences with participants and Ventana/Roche, it was decided to modify and relent the assessment criteria accepting larger negative areas in the individual tissue cores providing these still reliably could be scored concerning HER2/chr17 ratio. However, a slide with large negative areas was not compatible with an optimal assessment mark and was downgraded to good providing an otherwise optimal result being observed. The negative areas observed are random artefacts especially observed for the Ventana/Roche HER2 BRISH systems and an artefact recognized by both Ventana/Roche, NordiQC and the participants. In daily practice the end-user decides if samples with false negative areas can be scored or needs to be retested. The decision to relent the criteria was also based on the fact, that virtually all participants now use same or similar protocol settings for HER2 BRISH being locked by the vendor and thus not possible to optimize these further.

In both this and in concordance to the four previous assessments (runs H23, H24, H25 and H26), the combined GPA assay (VENTANA HER2 Dual ISH 800-6043 + HER2 IHC) was found less successful compared to the "standard" VENTANA HER2 Dual ISH assay. The "standard" VENTANA HER2 Dual ISH assay provided a pass rate of 88%, whereas the GPA assay gave a pass rate of 69%. The GPA assay was more successful compared to run H26 with a pass rate of only 13%, but still inferior to the "standard" assay

omitting HER2 IHC. The insufficient results were characterized by a successful IHC test for HER2 and as such showing a distinct and strong 3+ IHC membranous reaction of the neoplastic cells in the tumor tissue core no. 4, but showed only scattered cells displaying HER2 gene signals despite being highly amplified (Her2/chr17 ratio of 6.93-7.96 and > 9 HER2 signals pr cell). The central protocol settings e.g. HIER time/temp., HIER buffers and proteolysis, reported for the GPA assay, were similar to the settings reported for the single use VENTANA HER2 Dual ISH 800-6043 assay and thus not possible to identify any protocol parameters causing the very low pass rate in these three runs. However, one plausible explanation might be related to the strong 3+ HER2 IHC reaction in the tumour cells obscuring the penetration of HER2/chr 17 BRISH probes and/or enzymatic visualization of these. The result was downgraded as the HER2/chr17 signals could neither be evaluated in the tumour cells or the intermingling normal cells as expected and both cellular entities could be identified in the "standard" VENTANA HER2 Dual ISH without adding IHC to the protocol.

### Optimal protocol settings: Two-colour HER2 systems

152 laboratories used the **VENTANA Dual ISH system 800-6043** (Ventana/Roche).

Optimal demonstration of HER2 BRISH using this assay was typically based on the vendor recommended protocol settings based on a 2-step Heat Induced Epitope Retrieval (HIER) procedure using Cell Conditioning 1 (CC1) at 84°C followed by Cell Conditioning 2 (CC2) at 82°C for a total of 40 min. and subsequent proteolysis in ISH Protease 3 or Protease 3 for 8-20 min. at 36-37°C. The HER2 and chr17 probe cocktail being applied for 60 min. at 44°C following a denaturation step at 80°C for 8 min. – both steps and parameters are fixed by the vendor.

Among the laboratories reporting these protocol settings a pass rate of 87% (55 of 63) was obtained, 44% (28 of 63) being optimal.

16 laboratories used the **VENTANA Dual ISH system 800-6043** (Ventana/Roche) in combination with immunohistochemical demonstration for **HER2 PATHWAY®** (Ventana/Roche). The optimal result using this GPA assay, was based on HIER in CC1 or CC2 for 40-56 min. and subsequent proteolysis in ISH Protease 3 for 20 min. at 36-37°C.

Among the laboratories reporting these protocol settings a pass rate of 62% (8 of 13) was obtained, 31% (4 of 13) being optimal.

4 laboratories used the **ZytoDot® 2C system C-3022/C-3032** (ZytoVision).

2 protocols provided an optimal demonstration of HER2 BRISH and was based on HIER in EDTA, PT-0002-500 for 15 min. at 95°C, proteolysis in pepsin for 3-7,5 min. at 37°C, hybridization at 37°C for 12-24 hours following a denaturation at 75°C for 6-10 min. and visualization with the ZytoVision detection kit C-3022.

### HER2 ISH interpretation and scoring consensus

Table 3. **NordiQC ISH amplification data\***

	<b>NordiQC ISH HER2/chr17 ratio</b>	<b>NordiQC ISH HER2 copies</b>	<b>NordiQC HER2 amplification status</b>
1. Breast carcinoma	2.44-3	>6 (6.1-6.9)	Amplified
2. Breast carcinoma	1.19-1.66	<4 (2.15-3.65)	Non-amplified
3. Breast carcinoma	1.33-1.55	<3 (2.65-2.95)	Non-amplified
4. Breast carcinoma	6.93-7.96	>9 (9.95-11.45)	Amplified
5. Breast carcinoma	0.93-1.03	<2 (1.9-1.95)	Non-amplified

\* data from two NordiQC reference laboratory.

Table 4 shows the ISH assays used by the participants and concordance level to the NordiQC data observed. No technical evaluation of FISH protocols was performed. It has to be emphasized that it was not possible to identify the cause of an aberrant interpretation of the HER2 status whether this was related to the technical performance of the FISH assay or the interpretation by the observer(s).

Table 4. **ISH assays used and level of consensus HER2 status to NordiQC reference data, H27**


<b>BRISH</b>	n*	Vendor	Consensus	No consensus	Consensus rate
INFORM™ HER2 Dual ISH <b>780-4422/ 800-4422</b>	1	Ventana/Roche	1	0	-
VENTANA HER2 Dual ISH <b>800-6043</b>	128	Ventana/Roche	112	16	88%
VENTANA HER2 Dual ISH + IHC <b>800-6043 + HER2 IHC (GPA)</b>	13	Ventana/Roche	11	2	85%
ZytoDot® 2C <b>C-3022 / C-3032</b>	4	ZytoVision	3	1	-
ZytoDot® <b>C-3003</b>	1	ZytoVision	1	0	-
<b>FISH</b>					
PathVysion HER-2 DNA <b>6N4630 / 30-161060</b>	11	Abbott	11	0	100%
HER2 IQFISH <b>GM333</b>	6	Dako/Agilent	6	0	100%
HER2 AMP Probe <b>LPS001</b>	1	CytoCell	0	1	-
CytoTest <b>CT-PAC001</b>	1	CytoTest Inc	1	0	-
HER2 IQFISH <b>K5731</b>	14	Dako/Agilent	14	0	100%
SureFISH <b>G110104/G110993</b>	2	Dako/Agilent	2	0	-
PrimeFISH <b>17-012</b>	1	Diagen	1	0	-
HER2/CEN17 FISH probe <b>MF2001</b>	1	Fuzhou Maixin	1	0	-
BOND HER2 FISH system <b>TA9217</b>	9	Leica Biosystems	9	0	100%
FISH Kit <b>MAD-FISH-001</b>	2	Master Diagnostica	1	1	-
Rembrandt Her-2-C17 probe <b>C801P5206</b>	1	PanPath	1	0	-
ZytoLight <b>Z-2015 / Z-2020/ Z-2077</b>	17	ZytoVision	16	1	94%
ZytoMation ERBB2/CEN17 Dual Color FISH Probe <b>Z-2292</b>	4	ZytoVision	4	0	-
Total	217		195	22	
Proportion			90%	10%	

\*The number varies from Table 2. Not all participants have submitted a scoring sheet.


In total 217 of the 261 (83%) registered participants completed scoring sheets on the NordiQC website. These evaluations were compared to the HER2 ISH amplification status obtained by the NordiQC reference laboratories, summarized in Graph 2 and 3 (see page 6). For all laboratories performing either FISH or BRISH, the overall consensus rate was 90%. The consensus level for laboratories performing FISH was 96% (67 of 70 laboratories) and superior to the level of 87% for BRISH (128 of 147 laboratories). The level for FISH was on to level observed in previous assessments runs H24-H26, whereas the level was reduced for BRISH compared to previous runs.

Participants overall interpretation of amplification ratios and consensus rates are shown in Graph 2 and 3.

Graph 2: NordiQC HER2 ISH run H27: Participant interpretation of amplification status




Graph 3: NordiQC HER2 ISH run H27: Consensus depending on method




**Conclusion**


In this assessment an overall high pass rate of 85% was observed for BRISH. A technical optimal demonstration of HER2 BRISH was obtained by the widely used Ventana/Roche two-colour HER2 system **VENTANA HER2 Dual ISH 800-6043** and the modified GPA version of this assay. An optimal result could also be obtained by the Zytovision **ZytoDot® 2C C-3022/C-3032** system. Focusing on the technical quality of the HER2 BRISH assays an increase in the pass rate of 85% was observed compared to the level of 78 % in the previous run H26. In particular, in this assessment the modified GPA version of the VENTANA HER2 Dual ISH system performed with an improved pass rate at 69%, which was significantly better than the 13% pass rate seen in H26. The insufficient results were mainly caused by generally too weak or completely false negative results in one or more of the included tissue cores. In addition, also impaired morphology and more artefacts in combination characterized the insufficient results.




**Fig. 1a**  
Optimal demonstration of the HER2 gene status using the VENTANA HER2 Dual ISH kit cat. no. 800-6043, Ventana/Roche, of the breast carcinoma no. 1 with HER2 gene amplification:  
HER2/chr17 ratio 2.44 - 3, >6 HER2 copies\*.  
The HER2 genes are stained black and chr17 red.  
NordiQC and most participants interpreted this tumour as amplified.




**Fig. 1b**  
Optimal demonstration of the HER2 gene status using the VENTANA HER2 Dual ISH kit cat. no. 800-6043, Ventana/Roche, of the breast carcinoma no. 3 without HER2 gene amplification:  
HER2/chr17 ratio 1.33 - 1.55, <3 HER2 copies\*.  
The HER2 genes are stained black and chr17 red.  
The morphology is well preserved, and signals distinctively demonstrated.  
NordiQC and virtually all participants interpreted this tumour as non-amplified.




**Fig. 2a**  
Optimal demonstration of the HER2 gene status using the VENTANA HER2 Dual ISH kit cat. no. 800-6043, Ventana/Roche, of the breast carcinoma no. 4 with HER2 gene amplification:  
HER2/chr17 ratio 6.93 - 7.96, >9 HER2 copies\*.  
The HER2 genes are stained black and chr17 red.  
The signals are distinctively demonstrated in all the neoplastic cells.  
NordiQC and virtually all participants interpreted this tumour as amplified.




**Fig. 2b**  
Optimal demonstration of the HER2 gene status using the VENTANA HER2 Dual ISH kit cat. no. 800-6043, Ventana/Roche, of the breast carcinoma no. 5 without HER2 gene amplification:  
HER2/chr17 ratio 0.93 - 1.03, <2 HER2 copies\*.  
The HER2 genes are stained black and chr17 red.  
NordiQC and virtually all participants interpreted this tumour as non-amplified.




**Fig. 3a**  
Optimal demonstration of the HER2 gene status using the ZytoDot® 2C C-3022/C-3032, ZytoVision, of the breast carcinoma no. 5 without HER2 gene amplification: HER2/chr17 ratio 0.93- 1.3, <2 HER2 copies\*. The HER2 genes are stained green and chr17 red. NordiQC and virtually all participants also interpreted this tumour as non-amplified.



**Fig. 3b**  
Optimal demonstration of the HER2 gene status using the ZytoDot® 2C C-3022/C-3032, ZytoVision, of the breast carcinoma no. 1 with HER2 gene amplification: HER2/chr17 ratio 2.44 - 3, >6 HER2 copies\*. The HER2 genes are stained green and chr17 red. NordiQC and virtually all participants also interpreted this tumour as amplified.



**Fig. 4a**  
Sufficient demonstration of the HER2 gene status using the VENTANA HER2 Dual ISH kit cat. no. 800-6043, Ventana/Roche, in combination with HER2 IHC using PATHWAY, Ventana/Roche, of the breast carcinoma 3 without HER2 gene amplification: HER2/chr17 ratio 1.33 - 1.55, <3 HER2 copies\*. The gene protein assay (GPA) labels the HER2 genes black, chr17 red and HER2 protein brown. The IHC level is interpreted as 1+ and the GPA assay visualizes the HER2 protein expression and the HER2/chr17 gene status simultaneously. Despite some cells are lacking signals, the HER2 gene status can be established. The participant interpreted this tumour as non-amplified being concordant to the status determined by NordiQC and virtually all participants



**Fig. 4b**  
Insufficient demonstration of the HER2 gene status using the VENTANA HER2 Dual ISH kit cat. no. 800-6043, Ventana/Roche, in combination with HER2 IHC using PATHWAY, Ventana/Roche, of the breast carcinoma no. 4 with HER2 gene amplification: HER2/chr17 ratio 6.93 - 7.96, >9 HER2 copies\*. The gene protein assay (GPA) labels the HER2 genes black, chr17 red and HER2 protein brown. The IHC level is interpreted as 3+ but the vast majority of both neoplastic and stromal cells are totally negative concerning HER2 and Chr 17 signals and thus cannot reliably be scored. The ISH protocol applied was similar to successful ISH protocols and the aberrant staining reaction was most likely caused by the IHC DAB chromogen deposition in the GPA assay hindering penetration of the probes in the cells.






Fig. 5a  
 Insufficient demonstration of the HER2 gene status using the VENTANA HER2 Dual ISH kit cat. no. 800-6043, Ventana/Roche, of the breast carcinoma no. 4 with HER2 gene amplification: HER2/chr17 ratio 6.93 – 7.96, >9 HER2 copies\*. Virtually all cells are negative for both HER2 and chr17 signals and the HER2 gene status cannot be determined. This aberrant staining reaction / “negative spot artefact” was most likely caused by a technical issue during the staining process in the BenchMark instrument. Compare with Fig. 2a – same tumour and protocol, with an optimal result.

\* Range of data from FISH and BRISH performed in two NordiQC reference laboratories.




Fig. 5b  
 Insufficient demonstration of the HER2 gene status using the VENTANA HER2 Dual ISH kit cat. no. 800-6043, Ventana/Roche, of the breast carcinoma no. 5 without HER2 gene amplification: HER2/chr17 ratio 0.93 – 1.03, <2 HER2 copies\*. The HER2 genes are stained black and chr17 red. Virtually all cells are negative for chr17 signals and the HER2 gene status cannot reliably be determined. This aberrant reaction most likely was caused by a technical problem during the staining process in the BenchMark instrument or excessive retrieval which can cause missing chr17 signals. Figs. 1b – same tumour.

TB/SN/LE 25.04.2025