

Recommended protocol for TdT

Obtained in run 52

08 Dec 2017

Immunostainer

Type: Intellipath Biocare medical

Primary antibody

Clone: SEN28
Producer: Leica/novocastra
Product no. / lot no.: NCL-L-TdT-339 / 6039147
Diluent: Davinci green
Dilution factor: 1:50
Incubation time / temperature: 30 min. / 20°C

Epitope retrieval, HIER

Device: Pressure Cooker
Buffer: Biocare medical Borg
Heating time at max. temp.: 15 min.
Maximum heating temp.: 110°C

Visualization system

Producer: Biocare medicare
Product / no: Mach4 universal HRP polymer / BRI 4012L
Linker: None
Incubation time polymer: 10 min.
Incubation temperature: 20°C

Chromogen

Producer: Biocare medical
Product / no: Mach4 universal DAB / BRI 4012L
Incubation time / temperature: 5 min. / 20°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.