

Recommended protocol for BSAP

Obtained in run 53

16 Feb 2018

Immunostainer

Type: DBS Montage 360 system

Primary antibody

Clone: BV6
Producer: DBS
Product no. / lot no.: RMAB027 / I139
Diluent: Primary Antibody Diluent
Dilution factor: 1:10
Incubation time / temperature: 30 min. / 20°C

Epitope retrieval, HIER

Device: Pressure Cooker
Buffer: DBS DBS Montage Citrate Antigen Retrieval Solution
Heating time at max. temp.: 15 min.
Maximum heating temp.: 105°C

Visualization system

Producer: DBS
Product / no: Montage PolyVue Plus Auto Detection System / PVP100-A
Linker: Mouse and Rabbit Link
Incubation time linker: 10 min.
Incubation time polymer: 10 min.
Incubation temperature: 20°C

Chromogen

Producer: DBS
Product / no: Stable DAB/Plus / K047
Incubation time / temperature: 5 min. / 20°C
Enhancement: CuSO₄

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.