

Recommended protocol for p63

Obtained in run 48

19 Aug 2016

Immunostainer

Type: Gene Tech Genestainer

Primary antibody

Clone: 4A4
Producer: Zeta
Product no. / lot no.: Z2003L / 160112
Diluent: Antibody diluent
Dilution factor: 1:120
Incubation time / temperature: 50 min. / 25°C

Epitope retrieval, HIER

Device: Pressure Cooker
Buffer: Std Tris-EDTA / EGTA pH 9
Heating time at max. temp.: 3 min.
Maximum heating temp.: 110°C

Visualization system

Producer: Gene Tech
Product / no: GTVision / GK6005
Linker: None
Incubation time polymer: 20 min.
Incubation temperature: 25°C

Chromogen

Producer: Gene Tech
Product / no: DAB Detection Kit / GK6005
Incubation time / temperature: 5 min. / 25°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.