

## **Nordic Immunohistochemical Quality Control**

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

## **Recommended protocol for DOG1**

## Obtained in run

11 Sep 2015

**Immunostainer** 

Type: Leica BOND III

**Primary antibody** 

Clone: K9

Producer: Leica/novocastra

Product no. / lot no.: NCL-L-DOG-1 / 6012161
Diluent: Bond Antibody Diluent

Diluent producer / no.: Leica / AR9352

Dilution factor: 1:100

Incubation time / temperature: 30 min. / 20°C

Epitope retrieval, HIER

Device: On Board / On Machine

Buffer: Leica Bond Epitope Retrieval Solution 2

Heating time at max. temp.: 20 min. Maximum heating temp.: 100°C

Visualization system

Producer: Leica

Product / no: Bond Refine / DS9800

Incubation time linker: 8 min.
Incubation time polymer: 8 min.
Incubation temperature: 20°C

Chromogen

Producer: Leica

Product / no: Bond Refine / DS9800

Incubation time / temperature: 10 min. / 20°C

Enhancement: None

## Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.