

## Recommended protocol for ASMA

Obtained in run 55

14 Dec 2018

### Immunostainer

Type: Ventana Benchmark Ultra

### Primary antibody

Clone: EP188  
Producer: Epitomics  
Product no. / lot no.: AC-0167 / EO062007  
Diluent: Antibody Diluent  
Diluent producer / no.: Dako / S2022  
Dilution factor: 1:100  
Incubation time / temperature: 32 min. / 36°C

### Epitope retrieval, HIER

Device: On Board / On Machine  
Buffer: Ventana Ultra CC1  
Heating time at max. temp.: 32 min.  
Maximum heating temp.: 95°C

### Epitope retrieval, proteolysis

Enzyme: Protease 3  
Enzyme producer / no.: Ventana / 760-2020  
Incubation time / temp.: 12 min. / 36°C

### Visualization system

Producer: Ventana  
Product / no.: OptiView DAB IHC Detection Kit / 760-700  
Incubation time linker: 8 min.  
Incubation time polymer: 8 min.  
Incubation temperature: 36°C

### Chromogen

Producer: Ventana  
Product / no.: OptiView DAB IHC Detection Kit / 760-700  
Incubation time / temperature: 8 min. / 36°C  
Enhancement: CuSO<sub>4</sub>

### Disclaimer:



## **Nordic Immunohistochemical Quality Control**

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.