

Recommended protocol for WT1

Obtained in run 55

18 Dec 2018

Immunostainer

Type: Dako Autostainer Link 48 +

Primary antibody

Clone: WT49
Producer: Leica/Novocastra
Product no. / lot no.: NCL-L-WT1-562 / 06059847
Diluent: Antibody Diluent
Dilution factor: 1:10
Incubation time / temperature: 20 min. / 20°C

Epitope retrieval, HIER

Device: PT-link / PT-module
Buffer:
Heating time at max. temp.: 20 min.
Maximum heating temp.: 97°C

Visualization system

Producer:
Product / no: /
Linker: Linker, Mouse
Incubation time linker: 15 min.
Incubation time polymer: 20 min.
Incubation temperature: 20°C

Chromogen

Producer:
Product / no: /
Incubation time / temperature: 10 min. / 20°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.