

Recommended protocol for ASMA

Obtained in run 55

12 Feb 2019

Immunostainer

Type: Ventana Benchmark Ultra

Primary antibody

Clone: EP188
Producer: Epitomics
Product no. / lot no.: AC-0167 / EO62901
Diluent: Antibody Diluent
Diluent producer / no.: Dako / S2022
Dilution factor: 1:200
Incubation time / temperature: 16 min. / 36°C

Epitope retrieval, HIER

Device: On Board / On Machine
Buffer: Ventana Ultra CC1
Heating time at max. temp.: 32 min.
Maximum heating temp.: 95°C

Epitope retrieval, proteolysis

Enzyme: Protease 2
Enzyme producer / no.: Ventana / 760-2019
Incubation time / temp.: 4 min. / 36°C

Visualization system

Producer: Ventana
Product / no.: OptiView DAB IHC Detection Kit / 760-700
Incubation time linker: 8 min.
Incubation time polymer: 8 min.
Amplifier: OptiView Amplification Kit
Incubation time amplifier: 4 min.
Incubation temperature: 36°C

Chromogen

Producer: Ventana
Product / no.: OptiView DAB IHC Detection Kit / 760-700
Incubation time / temperature: 8 min. / 36°C
Enhancement: CuSO₄



Nordic Immunohistochemical Quality Control

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.