

Recommended protocol for SOX10

Obtained in run 55

02 Jan 2019

Immunostainer

Type: Gene Tech Genestainer

Primary antibody

Clone: SDM2
Producer: Zeta
Product no. / lot no.: Z2293 / 1430087B
Diluent: Ab diluent
Dilution factor: 1:60
Incubation time / temperature: 50 min. / 25°C

Epitope retrieval, HIER

Device: Water bath
Buffer: Std Tris-EDTA / EGTA pH 9
Heating time at max. temp.: 15 min.
Maximum heating temp.: 95°C

Visualization system

Producer: Gene Tech
Product / no: GTVision / GK6007
Linker: None
Incubation time polymer: 20 min.
Incubation temperature: 25°C

Chromogen

Producer: Gene Tech
Product / no: DAB Detection Kit / GK6007
Incubation time / temperature: 5 min. / 25°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.