

## Recommended protocol for ASMA

Obtained in run 55

02 Jan 2019

### Immunostainer

Type: Biocare Intellipath

### Primary antibody

Clone: 1A4  
Producer: Dako  
Product no. / lot no.: M0851 / 20046913  
Diluent: Tris with BSA  
Dilution factor: 1:250  
Incubation time / temperature: 30 min. / 23°C

### Epitope retrieval, HIER

Device: Other  
Buffer: Std Tris-EDTA / EGTA pH 9  
Heating time at max. temp.: 20 min.  
Maximum heating temp.: 100°C

### Visualization system

Producer: Dako  
Product / no: EnVision+, Mouse / K4000/K4001  
Linker: None  
Incubation time polymer: 30 min.  
Incubation temperature: 23°C

### Chromogen

Producer: Biocare  
Product / no: Betazoid DAB / DS900MM  
Incubation time / temperature: 10 min. / 23°C  
Enhancement: None

### Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.