

Recommended protocol for ASMA

Obtained in run 55

02 Jan 2019

Immunostainer

Type: Biocare Intellipath

Primary antibody

Clone: 1A4
Producer: Dako
Product no. / lot no.: M0851 / 20046913
Diluent: Tris with BSA
Dilution factor: 1:250
Incubation time / temperature: 30 min. / 23°C

Epitope retrieval, HIER

Device: Other
Buffer: Std Tris-EDTA / EGTA pH 9
Heating time at max. temp.: 20 min.
Maximum heating temp.: 100°C

Visualization system

Producer: Dako
Product / no: EnVision+, Mouse / K4000/K4001
Linker: None
Incubation time polymer: 30 min.
Incubation temperature: 23°C

Chromogen

Producer: Biocare
Product / no: Betazoid DAB / DS900MM
Incubation time / temperature: 10 min. / 23°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.