

## **Nordic Immunohistochemical Quality Control**

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

## Recommended protocol for ASMA

Obtained in run 55

04 Jan 2019

**Immunostainer** 

Type: Dako Omnis

**Primary antibody** 

Clone: BS66

Producer: Nordic Biosite

Product no. / lot no.: BSH-7459 / BSH100-4ai

Diluent: Antibody Diluent
Diluent producer / no.: Dako / K8006

Dilution factor: 1:1500

Incubation time / temperature: 30 min. / 24°C

Epitope retrieval, HIER

Device: On Board / On Machine

Buffer: Dako Omnis Target Retrieval Solution, High pH

Heating time at max. temp.: 30 min. Maximum heating temp.: 97°C

Visualization system

Producer: Dako Omnis

Product / no: EnVision Flex / GV800/GV823

Linker: Mouse LINKER

Incubation time linker: 10 min.
Incubation time polymer: 30 min.
Incubation temperature: 24°C

Chromogen

Producer: Dako Omnis

Product / no: DAB+ Substrate Chromogen System / GV825

Incubation time / temperature: 10 min. / 24°C

Enhancement: CuSO4

## Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.