

Recommended protocol for CD117

Obtained in run 56

07 Mar 2019

Immunostainer

Type: Thermo Autostainer 36/48/72

Primary antibody

Clone: YR145
Producer: Immunologic
Product no. / lot no.: VWRKILM7683-C1 / 1521005M
Diluent: normal antibody diluent
Dilution factor: 1:40
Incubation time / temperature: 60 min. / 20°C

Epitope retrieval, HIER

Device: PT-link / PT-module
Buffer: Dako TRS High pH (3-1)
Heating time at max. temp.: 10 min.
Maximum heating temp.: 97°C

Visualization system

Producer: Dako
Product / no: EnVision FLEX+ / K8002/SM802
Linker: None
Incubation time polymer: 30 min.
Incubation temperature: 20°C

Chromogen

Producer: Dako
Product / no: DAB+ / DM827 / DM823 / K800
Incubation time / temperature: 10 min. / 20°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.