

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

Recommended protocol for CK-LMW

Obtained in run

20 Sep 2015

Immunostainer

Type:

Ventana Benchmark Ultra

Primary antibody

Clone:	EP17
Producer:	Epitomics
Product no. / lot no.:	AC-0007 / CH062406
Diluent:	Antibody Diluent
Diluent producer / no.:	Dako / S2022
Dilution factor:	1:100
Incubation time / temperature:	32 min. / 36°C

Visualization system

Producer:	Ventana
Product / no:	OptiView DAB IHC Detection Kit / 760-700
Incubation time linker:	8 min.
Incubation time polymer:	8 min.
Incubation temperature:	36°C

Chromogen

-	
Producer:	Ventana
Product / no:	OptiView DAB IHC Detection Kit / 760-700
Incubation time / temperature:	8 min. / 36°C
Enhancement:	CuSO4

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.