

## **Nordic Immunohistochemical Quality Control**

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

## Recommended protocol for ALK (lung)

## Obtained in run 57

19 Aug 2019

**Immunostainer** 

Type: Dako Omnis

**Primary antibody** 

Clone: OTI1A4
Producer: Origene

Product no. / lot no.: UM800118 / F004
Diluent: Renoir Red Diluent

Dilution factor: 1:1000

Incubation time / temperature: 30 min. / 36°C

Epitope retrieval, HIER

Device: On Board / On Machine

Buffer: Dako Omnis Target Retrieval Solution, High pH

Heating time at max. temp.: 24 min. Maximum heating temp.: 97°C

Visualization system

Producer: Dako Omnis

Product / no: EnVision Flex / GV800/GV823

Linker: Mouse LINKER

Incubation time linker: 10 min.
Incubation time polymer: 20 min.
Incubation temperature: 36°C

Chromogen

Producer: Dako Omnis

Product / no: DAB+ Substrate Chromogen System / GV825

Incubation time / temperature: 10 min. / 36°C

Enhancement: None

## Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.