

Recommended protocol for CK8/18

Obtained in run 57

07 Oct 2019

Immunostainer

Type: Thermo Autostainer 36/48/72

Primary antibody

Clone: BSR15
Producer: Nordic Biosite
Product no. / lot no.: bsh-4000 / BSH-1bf
Diluent: Antibody Diluent
Dilution factor: 1:100
Incubation time / temperature: 30 min. / 21°C

Epitope retrieval, HIER

Device: PT-link / PT-module
Buffer: Std Tris-EDTA / EGTA pH 9
Heating time at max. temp.: 20 min.
Maximum heating temp.: 98°C

Visualization system

Producer: Nordic Biosite
Product / no: HRP-Polymer anti-rabbit / KDB-10046
Linker: None
Incubation time polymer: 30 min.
Incubation temperature: 21°C

Chromogen

Producer: Nordic Biosite
Product / no: High contrast DAB / BCB-20032
Incubation time / temperature: 10 min. / 21°C
Enhancement: CuSO₄

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.