

Nordic Immunohistochemical Quality Control

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

Recommended protocol for TTF1

Obtained in run 58

17 Dec 2019

Immunostainer

Type: Leica BOND III

Primary antibody

Clone: SPT24 Producer: Leica

Product no. / lot no.: PA0364 / 64136

Format: Ready-To-Use (prediluted)

Incubation time / temperature: 15 min. / 20°C

Epitope retrieval, HIER

Device: On Board / On Machine

Buffer: Leica Bond Epitope Retrieval Solution 1

Heating time at max. temp.: 20 min. Maximum heating temp.: 97°C

Visualization system

Producer: Leica

Product / no: Bond Refine / DS9800

Incubation time linker: 8 min.
Incubation time polymer: 8 min.
Incubation temperature: 20°C

Chromogen

Producer: Leica

Product / no: Bond Refine / DS9800

Incubation time / temperature: 10 min. / 20°C

Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.