

Recommended protocol for SATB2

Obtained in run 58

20 Dec 2019

Immunostainer

Type: Ventana Benchmark XT

Primary antibody

Clone: EP281
Producer: PathnSitu
Product no. / lot no.: PR239 / R06239LA
Diluent: None
Dilution factor: 1:0
Incubation time / temperature: 32 min. / 37°C

Epitope retrieval, HIER

Device: On Board / On Machine
Buffer: Ventana CC1
Heating time at max. temp.: 60 min.
Maximum heating temp.: 95°C

Visualization system

Producer: Ventana
Product / no: UltraView Universal DAB Detection Kit / 760-500
Incubation time polymer: 8 min.
Incubation temperature: 37°C

Chromogen

Producer: Ventana
Product / no: ultraView Universal DAB Detection Kit / 760-500
Incubation time / temperature: 8 min. / 37°C
Enhancement: CuSO₄

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.