

## **Nordic Immunohistochemical Quality Control**

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

## Recommended protocol for ASMA

Obtained in run 59

03 Feb 2020

**Immunostainer** 

Type: Leica BOND III

**Primary antibody** 

Clone: ?sm-1

Producer: Leicabiosystems
Product no. / lot no.: PA0943 / 65028

Format: Ready-To-Use (prediluted)

Incubation time / temperature: 15 min. / 20°C

Visualization system

Producer: Leica

Product / no: Bond Refine / DS9800

Incubation time linker: 8 min.
Incubation time polymer: 8 min.
Incubation temperature: 20°C

Chromogen

Producer: Leica

Product / no: Bond Refine / DS9800

Incubation time / temperature: 10 min. / 20°C

Enhancement: CuSO4

## Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.