

Recommended protocol for p57

Obtained in run

23 Sep 2015

Immunostainer

Type: IntelliPath FLX, Biocare

Primary antibody

Clone: 25B2
Producer: Leica/novocastra
Product no. / lot no.: NCL-p57 / 6002171
Diluent: Da Vinci Green
Dilution factor: 1:100
Incubation time / temperature: 30 min. / 20°C

Epitope retrieval, HIER

Device: Pressure Cooker
Buffer: Biocare Diva Deloaker pH 6,2
Heating time at max. temp.: 15 min.
Maximum heating temp.: 110°C

Visualization system

Producer: Biocare
Product / no: Medical Mach4 / NS
Incubation time polymer: 30 min.
Incubation temperature: 20°C

Chromogen

Producer: Biocare
Product / no: DAB / Medical Mach4
Incubation time / temperature: 5 min. / 20°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.