

Nordic Immunohistochemical Quality Control

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

# **Recommended protocol for p57**

# Obtained in run

23 Sep 2015

#### Immunostainer

Type:

IntelliPath FLX, Biocare

## **Primary antibody**

Clone:
Producer:
Product no. / lot no.:
Diluent:
Diluent producer / no.:
Dilution factor:
Incubation time / temperature:

25B2 Leica/novocastra NCL-p57 / 6002171 Da Vinci Green Biocare / NS 1:100 30 min. / 20°C

## **Epitope retrieval, HIER**

Device:	Pressure Cooker
Buffer:	Biocare Diva Deloaker pH 6,2
Heating time at max. temp .:	15 min.
Maximum heating temp.:	110°C

#### Visualization system

Producer:	Biocare
Product / no:	Medical Mach4 / NS
Incubation time polymer:	30 min.
Incubation temperature:	20°C

# Chromogen

Producer: Product / no: Incubation time / temperature: Enhancement: Biocare DAB / Medical Mach4 5 min. / 20°C

#### **Disclaimer:**

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.

None