

Recommended protocol for CD45

Obtained in run 59

06 Feb 2020

Immunostainer

Type: DBS Montage 360 system

Primary antibody

Clone: 2B11/PD7/26
Producer: DBS
Product no. / lot no.: Mob040 / Mob040-pilot-Mfg
Diluent: Primary Antibody Diluent
Dilution factor: 1:100
Incubation time / temperature: 30 min. / 20°C

Epitope retrieval, HIER

Device: Pressure Cooker
Buffer: DBS DBS Montage EDTA Antigen Retrieval Solution
Heating time at max. temp.: 15 min.
Maximum heating temp.: 105°C

Visualization system

Producer: DBS
Product / no: PolyVue PlusMouse/Rabbit HRP / PVP100-AUTO
Linker: PolyVue Plus Mouse/Rabbit Enhancer
Incubation time linker: 10 min.
Incubation time polymer: 10 min.
Incubation temperature: 20°C

Chromogen

Producer: DBS
Product / no: Stable DAB/Plus / K047
Incubation time / temperature: 5 min. / 20°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.