

Recommended protocol for CD45

Obtained in run 59

18 Feb 2020

Immunostainer

Type: Dako Autostainer Link 48 +

Primary antibody

Clone: 2B11+PD7/26
Producer: Dako
Product no. / lot no.: M0701 / 20068050
Diluent: Antibody Diluent
Dilution factor: 1:100
Incubation time / temperature: 30 min. / 25°C

Epitope retrieval, HIER

Device: PT-link / PT-module
Buffer: Dako TRS High pH (3-1)
Heating time at max. temp.: 20 min.
Maximum heating temp.: 97°C

Visualization system

Producer: Dako
Product / no: EnVision FLEX+ / K8002/SM802
Linker: Linker, Mouse
Incubation time linker: 15 min.
Incubation time polymer: 20 min.
Incubation temperature: 25°C

Chromogen

Producer: Dako
Product / no: DAB+ / K5007 / K4065
Incubation time / temperature: 10 min. / 25°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.