

Recommended protocol for Ki67

Obtained in run

24 Sep 2015

Immunostainer

Type: Ventana Benchmark Ultra

Primary antibody

Clone: SP6
Producer: BioCare
Product no. / lot no.: CRM325 / 81711
Diluent: Van Gogh
Diluent producer / no.: Biocare /
Dilution factor: 1:30
Incubation time / temperature: 24 min. / 37°C

Epitope retrieval, HIER

Device: On Board / On Machine
Buffer: Ventana Ultra CC1
Heating time at max. temp.: 64 min.
Maximum heating temp.: 100°C

Visualization system

Producer: Ventana
Product / no: UltraView Universal DAB Detection Kit / 760-500
Incubation time polymer: 8 min.
Amplifier: Ultraview/iView Amplification Kit
Incubation time amplifier: 8 min.
Incubation temperature: 37°C

Chromogen

Producer: Ventana
Product / no: ultraView Universa DAB Detection Kit / 760-500
Incubation time / temperature: 8 min. / 37°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.