

Recommended protocol for SOX10

Obtained in run 60

14 Aug 2020

Immunostainer

Type: intelliPATH

Primary antibody

Clone: BC34
Producer: Biocare Medical
Product no. / lot no.: API 3099 AA or H / 043020A
Format: Ready-To-Use (prediluted)
Incubation time / temperature: 30 min. / 25°C

Epitope retrieval, HIER

Device: Pressure Cooker
Buffer: Biocare Medical Diva Decloaker
Heating time at max. temp.: 110 min.
Maximum heating temp.: 15°C

Visualization system

Producer: Biocare Medical
Product / no: MACH4 HRP polymer / MRH534
Linker: MACH4 Mouse probe
Incubation time linker: 10 min.
Incubation time polymer: 20 min.
Incubation temperature: 25°C

Chromogen

Producer: Biocare Medical
Product / no: intelliPATH DAB Chromogen / IP5010K
Incubation time / temperature: 5 min. / 25°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.