

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

# **Recommended protocol for CD68**

# Obtained in run 60

14 Aug 2020

# Immunostainer

Type:

Dako Omnis

## Primary antibody

Clone:	KP1
Producer:	Dako Agilent
Product no. / lot no.:	GA609 / -
Format:	Ready-To-Use (prediluted)
Incubation time / temperature:	20 min. / 32°C

## Epitope retrieval, HIER

Device:
Buffer:
Heating time at max. temp.:
Maximum heating temp.:

On Board / On Machine Dako Omnis Target Retrieval Solution, High pH 30 min. 97°C

### Visualization system

Producer: Product / no: Linker: Incubation time polymer: Incubation temperature:

### Chromogen

Producer: Product / no: Incubation time / temperature: Enhancement: Dako Omnis EnVision Flex / GV800/GV823 None 20 min. 32°C

Dako Omnis DAB+ Substrate Chromogen System / GV825 10 min. / 32°C None

### Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.