

Recommended protocol for MLA

Obtained in run 60

07 Oct 2020

Immunostainer

Type: intelliPATH FLX

Primary antibody

Clone: M2-7C10
Producer: Zytomed Systems
Product no. / lot no.: MSK056 / A0431
Diluent: Antibody Diluent
Dilution factor: 1:75
Incubation time / temperature: 45 min. / 23°C

Epitope retrieval, HIER

Device: Other
Buffer: Zytomed Systems HIER Citrate Buffer
Heating time at max. temp.: 8 min.
Maximum heating temp.: 110°C

Visualization system

Producer: Zytomed Systems
Product / no: HRP Polymer / POLHRP100
Linker: Postblock
Incubation time linker: 20 min.
Incubation time polymer: 30 min.
Incubation temperature: 23°C

Chromogen

Producer: Zytomed Systems
Product / no: DAB high contrast / DAB500plus
Incubation time / temperature: 10 min. / 23°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.