

Recommended protocol for CD99

Obtained in run

24 Sep 2015

Immunostainer

Type: Ventana Benchmark Ultra

Primary antibody

Clone: EPR3097Y
Producer: Epitomics
Product no. / lot no.: 2429-1 / YF032508
Diluent: 0.025M TBS pH 7.4
Diluent producer / no.: /
Dilution factor: 1:1500
Incubation time / temperature: 32 min. / 36°C

Epitope retrieval, HIER

Device: On Board / On Machine
Buffer: Ventana Ultra CC1
Heating time at max. temp.: 60 min.
Maximum heating temp.: 99°C

Visualization system

Producer: Ventana
Product / no: UltraView Universal DAB Detection Kit / 760-500
Incubation time polymer: 8 min.
Incubation temperature: 36°C

Chromogen

Producer: Ventana
Product / no: ultraView Universa DAB Detection Kit / 760-500
Incubation time / temperature: 8 min. / 36°C
Enhancement: CuSO4

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.