

## **Nordic Immunohistochemical Quality Control**

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

## Recommended protocol for MLA

## Obtained in run 60

28 Aug 2020

**Immunostainer** 

Type: Gene Tech Genestainer

**Primary antibody** 

Clone: EP43

Producer: Epitomics

Product no. / lot no.: AC-0041A / EP041603

Diluent: Ab diluent

Diluent producer / no.: Gene Tech / GT100910

Dilution factor: 1:50

Incubation time / temperature: 50 min. / 25°C

Epitope retrieval, HIER

Device: Water bath

Buffer: Std Tris-EDTA / EGTA pH 9

Heating time at max. temp.: 15 min. Maximum heating temp.: 95°C

Visualization system

Producer: Gene Tech

Product / no: GTVision / GK6007

Linker: None Incubation time polymer: 20 min. Incubation temperature: 25°C

Chromogen

Producer: Gene Tech

Product / no: DAB Detection Kit / GK6007

Incubation time / temperature: 5 min. / 25°C

Enhancement: None

## Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.