

## **Nordic Immunohistochemical Quality Control**

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

## Recommended protocol for p63

## Obtained in run 61

17 Dec 2020

**Immunostainer** 

Type: Dako Omnis

**Primary antibody** 

Clone: DAK-p63
Producer: Dako/Agilent

Product no. / lot no.: GA662 / 20083242

Format: Ready-To-Use (prediluted)

Incubation time / temperature: 25 min. / 32°C

Epitope retrieval, HIER

Device: On Board / On Machine

Buffer: Dako Omnis Target Retrieval Solution, Low pH

Heating time at max. temp.: 30 min. Maximum heating temp.: 97°C

Visualization system

Producer: Dako Omnis

Product / no: EnVision Flex / GV800/GV823

Linker: None Incubation time polymer: 20 min. Incubation temperature: 32°C

Chromogen

Producer: Dako Omnis

Product / no: DAB+ Substrate Chromogen System / GV825

Incubation time / temperature: 5 min. / 32°C

Enhancement: None

## Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.