

Recommended protocol for HER2 IHC

Obtained in run B31

05 Jan 2021

Immunostainer

Type: Dako Autostainer Link 48 +

Primary antibody

Clone: HercepTest (polyclonal)
Producer: Dako
Product no. / lot no.: SK001 / 20077982B
Format: Ready-To-Use (prediluted)
Incubation time / temperature: 30 min. / 20°C

Epitope retrieval, HIER

Device: Water bath
Buffer: Dako Herceptest
Heating time at max. temp.: 40 min.
Maximum heating temp.: 98°C

Visualization system

Producer: Dako
Product / no: Herceptest / SK001
Incubation time polymer: 30 min.
Incubation temperature: 20°C

Chromogen

Producer: Dako
Product / no: Herceptest / SK001
Incubation time / temperature: 10 min. / 20°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.