

## Recommended protocol for CD31

Obtained in run 62

10 Feb 2021

### Immunostainer

Type: Titan s

### Primary antibody

Clone: MX032  
Producer: Fuzhou Maixin  
Product no. / lot no.: MAB-0720 / 2101270720a  
Diluent: None  
Dilution factor: 1:0  
Incubation time / temperature: 30 min. / 26°C

### Epitope retrieval, HIER

Device: Other  
Buffer: Fuzhou Maixin DNS  
Heating time at max. temp.: 20 min.  
Maximum heating temp.: 99°C

### Visualization system

Producer: Fuzhou Maixin  
Product / no: Titan super / TT-0805  
Linker: super  
Incubation time linker: 13 min.  
Incubation time polymer: 13 min.  
Incubation temperature: 26°C

### Chromogen

Producer: Fuzhou Maixin  
Product / no: DAB / TT-0805  
Incubation time / temperature: 6 min. / 26°C  
Enhancement: None

### Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.