

## Recommended protocol for DES

Obtained in run

27 Sep 2015

### Immunostainer

Type: Autostainer 720, LabVision

### Primary antibody

Clone: Polyclonal  
Producer: Thermo Scientific  
Product no. / lot no.: RB-9014-P / 9014P1005A  
Diluent: Antibody Diluent  
Dilution factor: 1:800  
Incubation time / temperature: 60 min. / 20°C

### Epitope retrieval, HIER

Device: PT-link / PT-module  
Buffer: Std Citrate pH 6  
Heating time at max. temp.: 20 min.  
Maximum heating temp.: 100°C

### Visualization system

Producer: Immunologic  
Product / no: NS / DPVO-999HRP  
Linker: None  
Incubation time polymer: 30 min.  
Incubation temperature: 20°C

### Chromogen

Producer: Immunologic  
Product / no: DAB / BS04-999  
Incubation time / temperature: 8 min. / 20°C  
Enhancement: None

### Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.