

Nordic Immunohistochemical Quality Control

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

Recommended protocol for CK19

Obtained in run

27 Sep 2015

Immunostainer

Type: IntelliPath, Biocare

Primary antibody

Clone: Ks19.1 Producer: BioCare

Product no. / lot no.: CM 242 / 041510-2 Diluent: Antibody Diluent

Dilution factor: 1:200

Incubation time / temperature: 30 min. / 20°C

Epitope retrieval, HIER

Device: Pressure Cooker

Buffer: NS Borg
Heating time at max. temp.: 1 min.
Maximum heating temp.: 125°C

Visualization system

Producer: Biocare

Product / no: NS / M4BD534L

Incubation time polymer: 30 min. Incubation temperature: 20°C

Chromogen

Producer: Biocare

Product / no: DAB / M4BD534L Incubation time / temperature: 5 min. / 20°C

Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.