

Recommended protocol for VIM

Obtained in run

27 Sep 2015

Immunostainer

Type: Autostainer, LabVision

Primary antibody

Clone: SP20
Producer: NeoMarkers
Product no. / lot no.: RM-9120-S / 9120S1001E
Diluent: Antibody Diluent
Dilution factor: 1:400
Incubation time / temperature: 30 min. / 20°C

Epitope retrieval, HIER

Device: Water bath
Buffer: Dako TRS High pH
Heating time at max. temp.: 20 min.
Maximum heating temp.: 98°C

Visualization system

Producer: Dako
Product / no: EnVision+ Dual Link System-HRP / K4061
Incubation time polymer: 30 min.
Incubation temperature: 20°C

Chromogen

Producer: Dako
Product / no: DAB+ / K3467 / K3468
Incubation time / temperature: 10 min. / 20°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.