

Recommended protocol for CD30

Obtained in run 65

07 Feb 2022

Immunostainer

Type: Dako Autostainer Link 48 +

Primary antibody

Clone: Ber-H2
Producer: Dako
Product no. / lot no.: IS602 or IR602 / 41247667
Format: Ready-To-Use (prediluted)
Incubation time / temperature: 20 min. / 25°C

Epitope retrieval, HIER

Device: PT-link / PT-module
Buffer: Dako TRS Low pH (3-1)
Heating time at max. temp.: 20 min.
Maximum heating temp.: 95°C

Visualization system

Producer: Dako
Product / no: EnVision FLEX / K8000/SM802
Linker: Linker, Mouse
Incubation time linker: 15 min.
Incubation time polymer: 20 min.
Incubation temperature: 25°C

Chromogen

Producer:
Product / no: /
Incubation time / temperature: 10 min. / 25°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.