

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

# **Recommended protocol for BSAP**

### Obtained in run 66

22 Jul 2022

#### Immunostainer

Type:

Leica BOND III

## Primary antibody

Clone:
Producer:
Product no. / lot no.:
Diluent:
Diluent producer / no.:
Dilution factor:
Incubation time / temperature:

SP34 Cell Marque 312R-16 / 70411 Van Gogh Yellow Diluent bio care / PD902M 1:100 15 min. / 21°C

#### **Epitope retrieval, HIER**

Device:	On Board / On Machine
Buffer:	Leica Bond Epitope Retrieval Solution 1
Heating time at max. temp.:	20 min.
Maximum heating temp.:	100°C

#### Visualization system

Producer:	Leica
Product / no:	Bond Refine / DS9800
Incubation time linker:	8 min.
Incubation time polymer:	8 min.
Incubation temperature:	21°C

#### Chromogen

Producer:
Product / no:
Incubation time / temperature:
Enhancement:

Leica Bond Refine / DS9800 10 min. / 21°C None

#### Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.