

## Recommended protocol for MLH1

Obtained in run 49

31 Jan 2017

### Immunostainer

Type: Thermo Autostainer 36/48/72

### Primary antibody

Clone: BS29  
Producer: Nordic Biosite  
Product no. / lot no.: BSH-7208 / BSH1-4ad  
Diluent: Antibody Diluent  
Dilution factor: 1:100  
Incubation time / temperature: 30 min. / 21°C

### Epitope retrieval, HIER

Device: PT-link / PT-module  
Buffer: Std Tris-EDTA / EGTA pH 9  
Heating time at max. temp.: 20 min.  
Maximum heating temp.: 98°C

### Visualization system

Producer: Nordic Biosite  
Product / no: BioSite Histo Plus HRP Polymer anti-Mouse / KDB-10007  
Linker: None  
Incubation time polymer: 30 min.  
Incubation temperature: 21°C

### Chromogen

Producer: Nordic Biosite  
Product / no: DAB High Contrast / BCB-20032  
Incubation time / temperature: 10 min. / 21°C  
Enhancement: CuSO<sub>4</sub>

### Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.