

## Recommended protocol for SYP

Obtained in run 66

17 Aug 2022

### Immunostainer

Type: Dako Autostainer

### Primary antibody

Clone: DAK-SYNAP  
Producer: Dako/Agilent  
Product no. / lot no.: IR660 / IR 660/ 20046433  
Format: Ready-To-Use (prediluted)  
Incubation time / temperature: 20 min. / 23°C

### Epitope retrieval, HIER

Device: PT-link / PT-module  
Buffer: Dako TRS High pH (3-1)  
Heating time at max. temp.: 20 min.  
Maximum heating temp.: 97°C

### Visualization system

Producer: Dako  
Product / no: EnVision FLEX+ / K8002/SM802  
Linker: Linker, Mouse  
Incubation time linker: 15 min.  
Incubation time polymer: 20 min.  
Incubation temperature: 23°C

### Chromogen

Producer: Dako  
Product / no: DAB+ / DM827 / DM823 / K800  
Incubation time / temperature: 10 min. / 23°C  
Enhancement: None

### Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.