

Recommended protocol for CD163

Obtained in run

03 Jun 2016

Immunostainer

Type: Dako Autostainer

Primary antibody

Clone: EDHu-1
Producer: Serotec
Product no. / lot no.: MCA1853 / 1
Diluent: Diamond - Antibody Diluent
Dilution factor: 1:1
Incubation time / temperature: 2 min. / 1°C

Visualization system

Producer: Dako
Product / no: EnVision FLEX / K8010/DM822
Linker: None
Incubation time polymer: 1 min.
Incubation temperature: 1°C

Chromogen

Producer: Dako
Product / no: DAB+ / DM827 / DM823 / K800
Incubation time / temperature: 1 min. / 1°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.