

## Recommended protocol for CK8/18

Obtained in run 69

24 Aug 2023

### Immunostainer

Type: Gene Tech GeneStainer

### Primary antibody

Clone: 5D3  
Producer: Thermo  
Product no. / lot no.: MS-743-S / 743S2011A  
Diluent: Ab diluent  
Dilution factor: 1:50  
Incubation time / temperature: 50 min. / 25°C

### Epitope retrieval, HIER

Device: On Board / On Machine  
Buffer: Std Tris-EDTA / EGTA pH 9  
Heating time at max. temp.: 20 min.  
Maximum heating temp.: 95°C

### Visualization system

Producer: Gene Tech  
Product / no: GTVision / GK8007  
Linker: None  
Incubation time polymer: 20 min.  
Incubation temperature: 25°C

### Chromogen

Producer: Gene Tech  
Product / no: DAB / GK8007  
Incubation time / temperature: 5 min. / 25°C  
Enhancement: None

### Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.