

Nordic Immunohistochemical Quality Control

Institute of Pathology, Aalborg University Hospital, Ladegaardsgade 3, P.O.Box 561, DK-9100 Aalborg, Denmark

Recommended protocol for SOX10

Obtained in run

06 Jun 2016

Immunostainer

Type: IntelliPath

Primary antibody

Clone: BC34
Producer: Biocare

Product no. / lot no.:

Diluent:

Diluent producer / no.:

AC13099C / 010815

Antibody diluent

Nordic Biosite / NS

Dilution factor: 1:100

Incubation time / temperature: 30 min. / 20°C

Epitope retrieval, HIER

Device: Pressure Cooker

Buffer: Biocare Diva Decloaker pH 6,2

Heating time at max. temp.: 5 min.

Maximum heating temp.: 110°C

Visualization system

Producer: Biocare

Product / no: NS / M4U534L

Linker: NS

Incubation time linker: 10 min.
Incubation time polymer: 15 min.
Incubation temperature: 20°C

Chromogen

Producer: Biocare

Product / no: DAB / M4U534L Incubation time / temperature: 5 min. / 20°C

Enhancement: CuSO4

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.