

Recommended protocol for CD34

Obtained in run

06 Jun 2016

Immunostainer

Type: Autostainer, Labvision

Primary antibody

Clone: BS72
Producer: Nordic Biosite
Product no. / lot no.: BSH-2008-1 / BSH100-1ap
Diluent: Antibody Diluent
Dilution factor: 1:150
Incubation time / temperature: 30 min. / 20°C

Epitope retrieval, HIER

Device: PT-link / PT-module
Buffer: Std Tris-EDTA / EGTA pH 9
Heating time at max. temp.: 20 min.
Maximum heating temp.: 98°C

Visualization system

Producer: Nordic Biosite
Product / no: NS / KDB-10007
Incubation time polymer: 30 min.
Incubation temperature: 20°C

Chromogen

Producer: Nordic Biosite
Product / no: DAB / BCB-20032
Incubation time / temperature: 10 min. / 20°C
Enhancement: CuSO₄

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.