

Recommended protocol for CD8

Obtained in run 72

18 Jul 2024

Immunostainer

Type: abcarta FAIP-48T

Primary antibody

Clone: 815R4B2
Producer: abcarta
Product no. / lot no.: PA577 / PA240516C03
Diluent: Antibody Diluent
Diluent producer / no.: abcarta / PB001
Dilution factor: 1:0
Incubation time / temperature: 15 min. / 30°C

Epitope retrieval, HIER

Device: On Board / On Machine
Buffer: abcarta EDTA/ER2 pH9.0
Heating time at max. temp.: 20 min.
Maximum heating temp.: 100°C

Visualization system

Producer: abcarta
Product / no: HRP Polymer / PS300
Linker: Dual Linker
Incubation time linker: 8 min.
Incubation time polymer: 8 min.
Incubation temperature: 30°C

Chromogen

Producer: abcarta
Product / no: DAB / PS300
Incubation time / temperature: 5 min. / 30°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.