

Recommended protocol for CD8

Obtained in run 72

02 Aug 2024

Immunostainer

Type: Titan

Primary antibody

Clone: MX117
Producer: Fuzhou Maixin
Product no. / lot no.: MAB-1031 / 2403201031a
Diluent: None
Diluent producer / no.: (None) / None
Dilution factor: 1:0
Incubation time / temperature: 30 min. / 26°C

Epitope retrieval, HIER

Device: On Board / On Machine
Buffer: Fuzhou Maixin DNS
Heating time at max. temp.: 20 min.
Maximum heating temp.: 99°C

Visualization system

Producer: Fuzhou Maixin
Product / no: Titan Super / TT-0805
Linker: Super
Incubation time linker: 10 min.
Incubation time polymer: 10 min.
Incubation temperature: 26°C

Chromogen

Producer: Fuzhou Maixin
Product / no: DAB / TT-0805
Incubation time / temperature: 6 min. / 26°C
Enhancement: None

Disclaimer:

NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.