Nordic Immunohistochemical Quality Control
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

Recommended protocol for Bcl-2

Obtained in run 73

26 Nov 2024

Immunostainer

Type:	Ventana Benchmark Ultra

Primary antibody

Clone:	124
Producer:	Ventana Roche
Product no. / lot no.:	790-4464 / K28199
Format:	Ready-To-Use (prediluted)
Incubation time / temperature:	16 min. / 36°C

Epitope retrieval, HIER

Device:	On Board / On Machine
Buffer:	Ventana Ultra CC1
Heating time at max. temp.:	64 min.
Maximum heating temp.:	95°C

Visualization system

Producer:	Ventana
Product / no:	UltraView Universal DAB Detection Kit / 760-500
Incubation time polymer:	8 min.
Amplifier:	Ultraview/iView Amplification Kit
Incubation time amplifier:	4+4 min.
Incubation temperature:	36°C

Chromogen

Producer:	Ventana
Product / no:	ultraView Universal DAB Detection Kit / 760-500
Incubation time / temperature:	8 min. / 36°C
Enhancement:	CuSO4

Disclaimer:
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.