**Nordic Immunohistochemical Quality Control**
Department of Pathology, Aalborg University Hospital, Hospitalsbyen 5, DK-9260 Gistrup, Denmark

# Recommended protocol for Bcl-2

Obtained in run 73

03 Dec 2024

**Immunostainer**

| | |
|---|---|
| Type: | Leica BOND III |

**Primary antibody**

| | |
|---|---|
| Clone: | ZR130 |
| Producer: | Zeta Corporation |
| Product no. / lot no.: | Z2682RT / ZR11102021A-5 |
| Diluent: | Zeta Antibody diluent |
| Diluent producer / no.: | Zeta Corporation / ZD19 |
| Dilution factor: | 1:300 |
| Incubation time / temperature: | 40 min. / 24°C |

**Epitope retrieval, HIER**

| | |
|---|---|
| Device: | On Board / On Machine |
| Buffer: | Leica Bond Epitope Retrieval Solution 2 |
| Heating time at max. temp.: | 20 min. |
| Maximum heating temp.: | 100°C |

**Visualization system**

| | |
|---|---|
| Producer: | Leica |
| Product / no: | Bond Refine / DS9800 |
| Incubation time linker: | 15 min. |
| Incubation time polymer: | 20 min. |
| Incubation temperature: | 24°C |

**Chromogen**

| | |
|---|---|
| Producer: | Leica |
| Product / no: | Bond Refine / DS9800 |
| Incubation time / temperature: | 10 min. / 24°C |
| Enhancement: | None |

**Disclaimer:**
NordiQC makes every attempt to provide accurate and up-to-date information, yet NordiQC does not make any claim or warranty regarding the accuracy of the provided information nor does it represent that the contents of the web site and protocols reflect the most recent developments in immunohistochemistry at any point in time.